

A comparative study of root canal shaping using protaper universal and protaper next rotary files in preclinical dental education

Gül Çelik^{Corresp., 1}, Feyza Özdemir Kısacık², Emir Yılmaz³, Arife Yaran⁴, İhsan Furkan Ertuğrul⁵, Hikmet Orhan⁶

¹ Faculty of Dentistry, Endodontics, Suleyman Demirel University, Isparta, Turkey

² Vefa Oral and Tooth Health Center, Afyonkarahisar, Turkey

³ Dentplus Oral and Dental Health Center, Bursa, Turkey

⁴ Araklı Bayram Halil State Hospital, Trabzon, Turkey

⁵ Faculty of Dentistry, Department of Endodontics, Pamukkale University, Denizli, Turkey

⁶ Faculty of Medicine, Biostatistics and Medical Informatics, Suleyman Demirel University, Isparta, Turkey

Corresponding Author: Gül Çelik
Email address: gulcelik@sdu.edu.tr

Background. One of Dentistry evolution has occurred in Endodontics due to the advancement of rotary techniques for root canal preparation and its subsequent incorporation into the teaching of Dentistry undergraduates. This research aimed to evaluate the shaping ability of the third-year dental students as their first experience in rotary instrumentation using ProTaper Universal (PTU) and ProTaper Next (PTN) (Dentsply Maillefer) rotary instruments in simulated curved canals. **Methods.** Forty students instrumented 200 simulated canals with a 40° curvature in resin blocks according to the manufacturer's instructions with PTU and 39 students 195 canals with PTN files. The canals were prepared at a speed of 300 rpm and using a 16:1 reduction hand-piece powered by an electric motor (Xsmart, Dentsply Maillefer). The final apical preparation was set to F2 for the PTU and X2 for the PTN group. The change in canal curvature was evaluated based on Schneider technique using the AutoCAD 2007 software on post-digital photographs. The incidence of instrument fracture and deformation, the incidence of ledge, the change in working length (WL), and the working time were noted. The data were analyzed with Student's t-test and Chi-Square test at a significance level of 0.05 using SPSS. **Results.** The PTN maintained the original canal curvature better, resulting in fewer fractures and ledges, and shaped the canals faster than the PTU ($P < 0.05$). The mean curves of the resin canals after the instrumentation for the PTU and PTN groups were $24.03^\circ \pm 3.14^\circ$ and $25.64^\circ \pm 2.72^\circ$, respectively. Thirty-three (17.4%) PTU and 18 (9.3%) PTN files fractured ($p < 0.05$). Nine (4.5%) PTU and 2 (2.6%) PTN deformed ($p > 0.05$). The change in WL after instrumentation was $0.97\text{mm} \pm 0.95\text{mm}$ in PTU and $0.96\text{mm} \pm 0.80\text{mm}$ in PTN ($p > 0.05$).

The mean times were $627s \pm 18 s$ for PTU and $379s \pm 18 s$ for PTN ($p < 0.000$). **Discussion.** PTN can be preferred in severe curved of the root canals in terms of maintenance of the original canal curvature, superior instrument fracture and fewer ledges. Even if short training before preparation provided an acceptable level of canal shaping for preclinical students, the use of NiTi rotary instruments should be included in the undergraduate dental curriculum, contributing to an increase in the quality of root canal shaping, and consequently to an improvement of clinical experience of students.

A COMPARATIVE STUDY OF ROOT CANAL SHAPING USING PROTAPER UNIVERSAL AND PROTAPER NEXT ROTARY FILES IN PRECLINICAL DENTAL EDUCATION

Gül ÇELİK¹, Feyza ÖZDEMİR KISACIK², Emir YILMAZ³, Arife YARAN⁴, İhsan Furkan ERTUĞRUL⁵, Hikmet Orhan⁶

¹ Faculty of Dentistry, Department of Endodontics, Suleyman Demirel University, Isparta, Turkey

² Vefa Oral and Tooth Health Center, Afyonkarahisar, Turkey

³ Dentplus Oral and Dental Health Center, Bursa, Turkey

⁴ Araklı Bayram Halil State Hospital, Trabzon, Turkey

⁵ Faculty of Dentistry, Department of Endodontics, Pamukkale University, Denizli, Turkey

⁶ Faculty of Medicine, Biostatistics and Medical Informatics, Suleyman Demirel University, Isparta, Turkey

Running Title: preparation in preclinical education

Corresponding author:

Gül Çelik

Department of Endodontics

Faculty of Dentistry

İzmir Demokrasi University,

İzmir, Turkey

Phone: +905324730823

Fax: +902462370607

Email:

gulcelik2016@gmail.com

A COMPARATIVE STUDY OF ROOT CANAL SHAPING USING PROTAPER UNIVERSAL AND PROTAPER NEXT ROTARY FILES IN PRECLINICAL DENTAL EDUCATION

Background. One of Dentistry evolution has occurred in Endodontics due to the advancement of rotary techniques for root canal preparation and its subsequent incorporation into the teaching of Dentistry undergraduates. This research aimed to evaluate the shaping ability of the third-year dental students as their first experience in rotary instrumentation using ProTaper Universal (PTU) and ProTaper Next (PTN) (Dentsply Maillefer) rotary instruments in simulated curved canals. **Methods.** Forty students instrumented 200 simulated canals with a 40° curvature in resin blocks according to the manufacturer's instructions with PTU and 39 students 195 canals with PTN files. The canals were prepared at a speed of 300 rpm and using a 16:1 reduction hand-piece powered by an electric motor (Xsmart, Dentsply Maillefer). The final apical preparation was set to F2 for the PTU and X2 for the PTN group. The change in canal curvature was evaluated based on Schneider technique using the AutoCAD 2007 software on post-digital photographs. The incidence of instrument fracture and deformation, the incidence of ledge, the change in working length (WL), and the working time were noted. The data were analyzed with Student's t-test and Chi-Square test at a

significance level of 0.05 using SPSS. **Results.** The PTN maintained the original canal curvature better, resulting in fewer fractures and ledges, and shaped the canals faster than the PTU ($P < 0.05$). The mean curves of the resin canals after the instrumentation for the PTU and PTN groups were $24.03^\circ \pm 3.14^\circ$ and $25.64^\circ \pm 2.72^\circ$, respectively. Thirty-three (17.4%) PTU and 18 (9.3%) PTN files fractured ($p < 0.05$). Nine (4.5%) PTU and 2 (2.6%) PTN deformed ($p > 0.05$). The change in WL after instrumentation was $0.97\text{mm} \pm 0.95\text{mm}$ in PTU and $0.96\text{mm} \pm 0.80\text{mm}$ in PTN ($p > 0.05$). The mean times were $627\text{s} \pm 18\text{s}$ for PTU and $379\text{s} \pm 18\text{s}$ for PTN ($p < 0.000$). **Discussion.** PTN can be preferred in severe curved of the root canals in terms of maintenance of the original canal curvature, superior instrument fracture and fewer ledges. Even if short training before preparation provided an acceptable level of canal shaping for preclinical students, the use of NiTi rotary instruments should be included in the undergraduate dental curriculum, contributing to an increase in the quality of root canal shaping, and consequently to an improvement of clinical experience of students.

A COMPARATIVE STUDY OF ROOT CANAL SHAPING USING PROTAPER UNIVERSAL AND PROTAPER

NEXT ROTARY FILES IN PRECLINICAL DENTAL

EDUCATION

INTRODUCTION

Cleaning and shaping of the root canal, which must be performed mechanically and biologically, is the most important procedure in root canal treatment. A variety of techniques and instruments have been developed to perform this procedure without generating undesirable clinical results, such as canal straightening, transportation, ledging, strip perforations, or instrument fractures. The European Society of Endodontics (ESE) has published current articles to lead the undergraduate curriculum in endodontics to improve dental students' theoretical and clinical education (*ESE, 2001*). ESE believes that undergraduate students should receive preclinical and clinical training to be able to successfully treat uncomplicated anterior premolar and molar teeth (*De Moor et al., 2013*). Endodontic education needs to be developed in undergraduate programs (*Dummer, 1991; Qualtrough & Dummer, 1997; Unal et al., 2012*) and the use of new techniques and devices that have proven to be positive contributions to root canal therapy should be included in undergraduate training programs (*Unal et al., 2011*).

The technical quality of root fillings, assessed radiographically, performed by undergraduate students using hand instruments was investigated in a meta-analysis (*Riberio et al., 2018*). The results revealed a low frequency (48.75%) of acceptable technical quality of root fillings and drew attention to the major procedural errors created

(*Riberio et al., 2018*). Many studies have been conducted to evaluate the performance of students in curved root canals (*Hänni et al., 2003; Arbab-Chirani & Vulcain 2004; Alrahabi, 2015; Unal et al., 2011*). Despite the multiple benefits of using NiTi rotary instruments, the step-back technique using stainless steel files is still a conventional teaching method in endodontic programs for undergraduate students in most countries (*Alrahabi, 2015*). Although students with no experience are reported to be successful in the use of rotary instruments (*Unal et al., 2012; Kwak et al., 2016; Brito-Júnior et al., 2014*), it is evident that errors can be minimized by the training to be taken in the curriculum. *Georgelin-Gurgel et al. (2008)* and *Alrahabi (2015)* reported that manual instrumentation is safer than rotary instrumentation in the hands of inexperienced students. They underlined that acquiring skill in the use of NiTi rotary instrumentation requires specific preclinical training to avert file breakage. *Martins et al. (2012)* suggested that the use of NiTi rotary instruments should be included in the undergraduate dental curriculum, as this would contribute to an increase in the number of patients assisted, and consequently, improve the students' clinical experience.

Some procedural errors have been minimized after the introduction of the NiTi alloy into endodontics (*Walia et al., 1988; Vaudt et al., 2009; Yang et al., 2011*). Nevertheless, the fracture of the NiTi rotary root canal instruments suddenly and without any warning creates a disappointment for the dentist (*Zuolo & Walton 1997; Arens et al., 2003; Ankrum et al., 2004*). The studies are underway to increase the resistance of the root canal instrument against breakage in order to overcome this problem. A recent novelty of the alloy is the so-called m-wire alloy. The M-Wire material has been shown

to have longer fatigue resistance and lower fracture risk than conventional NiTi alloy does (Johnson et al., 2008; Montenegro-Santillan et al., 2013).

The ProTaper Universal (PTU) system (Dentsply Maillefer, Ballaigues, Switzerland) is currently used by endodontists, and it is one of the most revolving instrument systems researchers have searched for the most (*Peters et al., 2003; Schäfer & Vlassis, 2004; Celik Unal et al., 2006; Yang et al., 2007; Alemam et al., 2017*). PTU is nickel titanium (NiTi) rotary system of instruments manufactured with progressive tapering over the length of the cutting blades, convex triangular cross-sections, and noncutting tips. In the PTU system, the file is produced in such a way that it does not cut the ends, and its cross-section is a convex triangle. The taper angle of the file is not constant. This angle increases parabolically starting from the end (*Hieawy et al., 2015*).

The design features of ProTaper Next (PTN, Dentsply Maillefer), made from the new M-Wire alloy, include variable tapers and rectangle sections with a remote center. The M-Wire material has been shown to have longer fatigue resistance and lower fracture risk than conventional NiTi alloy does (*Johnson et al., 2008; Montenegro-Santillan et al., 2013*). The number of instruments is similar to that of the PTU in terms of the order of use and each instrument. The PTU contains three files (SX, S1, and S2) for the preparation of the coronal and middle thirds and three files (F1, F2, and F3) for the preparation of the apical third. PTN consists of only three files (X1 is #17/.04, X2 is #25/.06, and X3 is #30/.075), close to the diameters of the files used in the apical third in the PTU system (*Pérez-Higueras et al., 2014*).

This study aimed to compare the shaping ability of PTU and PTN rotary instruments used by the third-year dental students inexperienced in any rotary instrumentation techniques in simulated curved canals of resin blocks.

METHODS

Seventy-nine third-year undergraduate dental students (2015–2016 intakes) in preclinical endodontics at the School of the Dentistry of the Suleyman Demirel University, Isparta, Turkey had three 50-min lectures of about the NiTi instruments, their physical properties, and the special constructional features of the files. The present study was based on data obtained from practical courses, which was compulsory for the students to attend. Therefore, there was no need to sign an acceptance form for participation. The results of this study did not affect the course notes of the students. The students instrumented 16 simulated canals in acrylic resin blocks in the 2014-2015 academic year and 12 extracted human teeth 2015-2016 in the preclinical dental education. The shaping technique was Modified Double Flared Technique using balanced force principle (Saunders & Saunders, 1992) with stainless steel instrument. The students attended a 2-h lecture on the use of PTU and PTN files two weeks before the end of the term. The study was started in the last week of the term. The students were divided into groups of 10, and an associated professor (G.Ç.) demonstrated the procedures of the shaping of a simulated canal according to the instructions of the manufacturer. A printed script with the manufacturer's step-by-step instructions for the PTU and PTN rotary systems was given to each student.

The students instrumented a total of 395 simulated resin canals (Endo Training Block 02 taper, REFA 0177; Dentsply Maillefer, CH-1338 Ballaigues, Switzerland). All simulated resin canals had an apical foramen of 0.15 mm, a taper of 0.02, and an angle of curvature of 40°. Of the 79 students, 40 students assigned randomly instrumented 200 canals using PTU files and 39 students instrumented 195 canals using PTN files. Each student instrumented five simulated canals using just one of the two systems and only one set instrument. .. The canals for each file system were instrumented to a working length of 16 mm (0.5 mm from the apex) at a speed of 300 rpm and a torque-control level of 2, using a 16:1 reduction handpiece powered by an electric motor (Xsmart, Dentsply Maillefer). Two mL of distilled water was used as an irrigant at each change of instrument. After the instrumentation, a final rinse with distilled water for 1 min was carried out. Each canal was dried using size 25 paper points.

The incidence of fractures in the blocks and instrument deformations It were recorded. The blocks which instrument fractured were not taken into account when calculating the curvature and WL change, and incidence of the ledge.. The working time was calculated starting through the insertion of the first file until the end of the instrumentation, including total active instrumentation, cleaning of the flutes of the instruments, and irrigation. The final WL of the canals was determined in mm following instrumentation of canals. An F2 PTU or X2 PTN file was inserted into the canal and its WL within the canal measured to the nearest 0.5 mm. The amount of change in WL was determined by subtracting the final length from 16 mm.. Canal straightening and ledge formation were assessed on digital images (Figure 1 and Figure 2). After the canal instrumentation, the images of the resin blocks were obtained with the help of a light

microscope camera (Zeiss Axioskop 2, Munich, Germany) to determine the canal curvature. The images were taken from each block in one direction (mimicking clinical conditions). In these images, canal curvatures were measured by a computer program (AutoCAD 2007) according to Schneider (1971). The change in canal curvature was determined by subtracting the value obtained after instrumentation from 40 degrees. All students were randomly divided into two groups as they are at the same educational level. Therefore, it was accepted that the only variable in the experimental groups was PTU and PTN files. Data were analyzed using SPSS software, version 10.0 (SPSS Inc., Chicago, USA). The level of significance was set at 5%. Student's t-test was used to evaluate differences in canal straightening, working time, and amount of change in WL. The chi-square test was used to evaluate differences in the incidences of instrument fracture, instrument deformation, while Yates corrected chi-square test was used ledge formation between the groups.

RESULTS

The PTN preserved the original canal curvature ($P < 0.000$) better, resulting in fewer fractures ($p < 0.05$) and ledges ($p < 0.000$), and shaped the canals faster ($P < 0.000$) than the PTU. The mean curvatures of the canals after instrumentation with PTU and PTN were $24.03^\circ \pm 3.14^\circ$ and $25.64^\circ \pm 2.72^\circ$, respectively ($P < 0.000$). The incidence of instrument fracture was 17.4% (33) for PTU and 9.32 % (18) for PTN ($p < 0.05$). The incidence of instrument deformation was 9 (4.5 %) for PTU and 2 (2.6 %) for PTN ($P > 0.05$). The PTN had fewer ledges than PTU ($P < 0.000$). Ledge formation was observed in 33 (18.8%) of the canals in the PTU group, while in 7 (4 %) in the PTN group ($P < 0.000$). There was no statistically significant difference in the preservation of

the WL ($P>0.05$). The mean change in WL after instrumentation was $0.97\text{mm} \pm 0.95\text{ mm}$ in the PTU group and $0.96 \pm 0.80\text{ mm}$ in the PTN group ($p > 0.05$) PTN shaped the canals more quickly than PTU ($P<0.000$). The mean time in instrumentation of the canals was $627\text{s} \pm 28\text{s}$ for PTU and $379\text{s} \pm 18\text{s}$ for PTN ($P<0.000$) (Table 1).

DISCUSSION

The ESE, Education and Scholarship Committee, encourages the use of proven new tools, techniques, and training resources in endodontics education (*Dummer et al., 2013*). Updated of the endodontic clinical curriculum is an essential step to having acceptable endodontic treatment, for both general practitioners and specialist dentists. Preclinical exercises are extremely important for the acquisition of this skill (De Moor et al, 2013; Georgelin-Gurgel et al., 2008; Alrahabi, 2015).

This research presents the experiences of the third-year preclinical students in the Faculty of Dentistry of the University of Suleyman Demirel with rotary instruments following training on root canal preparation with hand instruments. Instrumentation techniques with engine-driven instruments are now almost indispensable for root canal treatment beyond popularization in root canal instrumentation. Along with hand instruments, which are indispensable, instrumentation with rotary instruments is also important. Hänni et al. (2003) showed that third-year students can use preclinical course profile 0.04 taper (Dentsply Maillefer, Ballagues, Switzerland) instruments in a satisfactory manner. Moreover, Arbab-Chirani & Vulcain (2004) in France and Abu-Tahun et al. (2014) in Jordan reported that there is a consensus on the need for the use

of rotary instruments in undergraduate education. Following this, the use of rotary instruments in the dentistry faculty was added to the curriculum in France. ; Kang et al. (2006) reported that the ProFile is the safest, best instrument for root canal shaping for students and beginners. Peru et al. (2006) reported that curved root canals prepared by non-experienced undergraduates with rotary instruments exhibit less procedural errors and require less time than those generated using hand instruments. In addition, Tu et al. (2008) showed that undergraduate students could generate more successful root canal shaping with ProTaper rotary instruments than they can with hand ProTaper instruments. Sonntag et al. (2008) found that preclinical endodontics training varied significantly according to the curriculum, instructor, and course content, and they announced that, as of 2008, most (63%) faculties in Germany taught root canal instrumentation with rotary Ni-Ti instruments. Unal et al. (2012) reported that third-year students with no experience with rotating instruments were successful in molar root canal preparations after 2 h of theoretical education about rotary instruments. Brito-júnior et al. (2014) reported that undergraduate students produced lower apical transportation in curved canals with F1 and F2 PTU files. Furthermore, Bruno et al. (2016) stated that the low fracture rates observed in their study indicated that the examined instruments can be used safely by students.

In the literature, there are many different experiments to evaluate the shaping and cleaning of the root canal. Artificial resin models are preferred in many studies because having a standard canal length, curvature, and form (Kum et al., 2000; Unal et al., 2009; Unal et al., 2012). The μ CT is emerging in several endodontic research facilities as a nondestructive and accurate method to analyze canal geometry and the relative effects

of shaping techniques (Peters et al., 2003; Peru et al., 2006). As the cost of the μ CT assessment method was high, our study was carried out in acrylic resin blocks. Although this experimental model does not fully reflect the morphology of real human teeth, it can give an idea of relation to relation to the performance of the root canal instrument.

The most frequently encountered case for procedural errors during the preparation of curved root canals is the straightening of the root canals instrument fracture and deformation, and ledge (Nagy et al. 1997). Some of the reasons for these procedural errors are the instrument type and dimension, the type of alloy and the canal curve of before the instrumentation (Parashos & Messer, 2006; Lambrianidis, 2009). Further, the test material in which the instrumentation is carried out may also affect the test results (Alrahabi & Zafar, 2018). In our study, PTN (25.64°) maintained original canal curvature better than PTU (24.03°) ($P < 0.000$). The amount of straightening was found to be higher compared with Unal et al.'s (2012) study. The researchers selected the F1 PTU

file as the master apical file. The smaller the size of this file than the F2 PTU file, the less straightening it may have caused. Besides, the shaping ability of the files was examined in acrylic resin blocks in our study, while Unal et al. (2012) tested in extracted human teeth. Alrahabi & Zafar (2018) reported that the files show different performance in the extracted human teeth and simulated resin canals.

The second of the most frustrating procedural errors encountered during the instrumentation of curved root canals is the abrupt fracturing and deforming of the instruments. Applying enough force to metal causes the shape of the material to

change, and this shape change is called deformation. After the force is removed, the permanent shape change that cannot be returned to its original state is called plastic deformation (*Thompson 2000*). The opening in the grooves manifests the permanent deformation. In this study, deformation was detected under a microscope. The incidence rate of deformation was 4.5% (9) for PTU and 2.6% (2) for PTN ($p > 0.05$).

The incidence rates of fracture were 17.4% for PTU and 9.3 % for PTN ($p < 0.05$) in our study. *Vieira et al. (2008)* reported that unexpected instrument fractures in the automatic systems were closely related to the operator's experience. On contrary, it was reported that inexperienced students could shape the root canal as satisfactory (*Hänni et al., 2003; Peru et al., 2006; Unal et al., 2012; Brito-Júnior et al., 2014*). *Bruno et al. (2016)* announced that a total of 30 file fractures were noted during the study period; thus, fractures occurred in 0.37% of total file uses and 2.98% of all works. The authors attributed a low rate of fractures to some factors. These factors are the use of the F3 ProTaper, the experience of the operator, and supervised treatment by expert endodontists. Also, *Bruno et al. (2016)* designed this study under clinical circumstances. Therefore, since the study included all tooth types, the average curvature of the teeth could be assumed to be less than 40 degrees. Besides, working on more curved canals and acrylic resin blocks instead of real teeth could be increased the numbers of instrument fractures and deformations. One of the disadvantages of the use of rotary canal instruments in resin blocks is that they heat the resin while the instruments are in operation, and this heat causes the resin to soften. This heat can also cause the canal instrument to bind to and become entangled in the softened resin, thereby breaking the instrument (*Baumann & Roth, 1999; Kum et al., 2000*). The choice of acrylic resin

blocks for standardization is also a limitation of this study. Nevertheless, in a study by Alemann et al. (2017) comparing the shaping abilities of the PTU and PTN instruments in S-shaped canals, the students were more successful. This difference may be due to their use of aqueous lubricant during the preparation. Lubrication usually helps prevent fracture of the instrument during root canal preparation by helping the instrument work in the canal without being forced (Zehnder 2006). We did not use lubricants in our study; we only used distilled water for irrigation. An aqueous lubricant has been shown to be more effective than a gel-type lubricant. A lubricant may exert a physical effect by floating debris away from the rotating instrument. In addition, chemical additives may act on the root canal dentin to facilitate instrumentation (Peters et al., 2005; Boessler et al., 2007). In our study, this difference between PTU and PTN in terms of instrument fracture and instrument deformation can be due to the different levels of flexibility of the instruments. While PTU is produced from a conventional NiTi alloy, PTN is produced from the more flexible M-Wire. Furthermore, The PTN instruments have an innovative off centered rectangular cross-section that gives the file a snake-like swaggering movement as it progresses through the root canal. The manufacturer claims that this asymmetric rotary motion of ProTaper Next allows achieving fully tapered canals with fewer numbers of files (Tulsa Dental Specialties 2017). In addition, the X2 of PTN has an apical cone of 0.06, while PTU's F2 instrument exhibits a constant taper from D1 to D3 (0.08) (Alemam et al., 2017).

A ledge is defined as a deviation from the original canal curvature within the apical third which creates or starts to create a new canal at a tangent to the original canal (Nagy et al. 1997). In our study, ledges occurred in 19% of the canals prepared with

PTU, whereas ledges formed in 4% of PTN cases. One of the reasons why PTN files produce fewer ledges than PTU files can be the superior properties of the m-wire alloy. The alloy structure and cross-sectional shape of the PTU system's F1 and F2 instruments will result in a hardened instrument that can cause more ledges to occur than arise in PTN. Furthermore, Alrahabi & Zafar (2018) considered contributing to the snake-like, swaggering movements of the files during advancement into the root canal. This movement serves to minimize the engagement between the file and dentin. Reduced engagement limits any undesirable taper lock, the screw effect, and the torque on any given file (Haapasalo & Shen (2013).

CONCLUSIONS

Within the limitations of this study, the PTN can be preferred in severe curved of the root canals in terms of maintenance of the original canal curvature, superior fracture, and fewer ledges. Even if short training before preparation provides an acceptable level of canal shaping for preclinical students, the use of NiTi rotary instruments should be included in the undergraduate dental curriculum, contributing to an increase in the quality of root canal shaping, and consequently the improvement of clinical experience of students.

ACKNOWLEDGEMENTS

The authors thank Dentsply® for the root canal files and resin blocks used in this study.

REFERENCES

Abu-Tahun I, Al-Rabab'ah MA, Hammad M, Khraisat A. 2014. Technical quality of root canal treatment of posterior teeth after rotary or hand preparation by fifth year undergraduate students, The University of Jordan. *Australian Endodontic Journal*. 40:123-30. DOI 10.1111/aej.12069.

Arbab-Chirani R, Vulcain JM. 2004. Undergraduate teaching and clinical use of rotary nickel-titanium endodontic instruments: a survey of French dental schools. *International Endodontic Journal*. 37:320-4. DOI 10.1111/j.0143-2885.2004.00805.x.

Alemam AAH, Dummer PMH, Farnell DJJ. 2017. A Comparative Study of ProTaper Universal and ProTaper Next Used by Undergraduate Students to Prepare Root Canals. *Journal of Endodontics*. 43:1364. DOI 10.1016/j.joen.2017.03.038.

Alrahabi M. 2015. Comparative study of root-canal shaping with stainless steel and rotary NiTi files performed by preclinical dental students. *Technol Health Care*. 23):257-65. DOI 10.3233/THC-150895.

Alrahabi M, Zafar MS. 2018. Assessment of apical transportation caused by nickel-titanium rotary systems with full rotation and reciprocating movements using extracted teeth and resin blocks with simulated root canals: A comparative study. *Nigerian Journal of Clinical Practice*. 21:772-777. DOI 10.4103/njcp.njcp_200_17.

Ankrum MT, Hartwell GR, Truitt JE. 2004. K3 Endo, ProTaper, and ProFile systems: breakage and distortion in severely curved roots of molars. *Journal of Endodontics*. 30:234-7. DOI 10.1097/00004770-200404000-00013

Arens F C, Hoen M M, Steiman H R, Dietz G C. 2003. Evaluation of single-use rotary nickel-titanium instruments. *Journal of Endodontics*. 29: 664–666.

Baumann MA, Roth A. 1999. Effect of experience on quality of canal preparation with rotary nickel-titanium files. *Oral Surgery, Oral Medicine, Oral Pathology, Oral Radiology, and Endodontology*. 88:714–8. DOI 10.1016/S1079-2104(99)70015-6.

Boessler C, Peters OA, Zehnder M. 2007. Impact of lubricant parameters on rotary instrument torque and force. *Journal of Endodontics*. 33:280-3. DOI 10.1016/j.joen.2006.11.007

Brito-Júnior M, Faria-E-Silva AL, Camilo CC, Pereira RD, Braga NM, Sousa-Neto MD. 2014. Apical transportation associated with ProTaper® Universal F1, F2 and F3 instruments in curved canals prepared by undergraduate students. *Journal of Applied Oral Science*. 22: 98–102. DOI 10.1590/1678-775720130464.

Bruno FA, Nunes E, Horta MC, da Fonseca AM, Silveira FF. 2016. Importance of rotary systems in dental care by undergraduate students in patients of a public health service of Belo Horizonte. *Journal of Clinical and Experimental Dentistry*. 8:e60-3. DOI 10.4317/jced.52663.

Celik Unal G, Kecici AD, Ureyen Kaya B. 2006. Comparison of the shaping abilities of two rotary and a hand instruments in simulated root canals. *Acta Odontologica Turcica*. 23:11–6.

De Moor R, Hülsmann M, Kirkevang LL, Tanalp J, Whitworth J. 2013. Undergraduate curriculum guidelines for endodontology. *International Endodontic Journal*. 46:1105-14. DOI 10.1111/iej.12186.

Dummer PM. 1991. Comparison of undergraduate endodontic teaching programmes in the United Kingdom and in some dental schools in Europe and the United States. International Endodontic Journal. 24: 169-77. DOI 10.1111/j.1365-2591.1991.tb00127.x.

European Society of Endodontology. 2001. Undergraduate curriculum guidelines for endodontology. International Endodontic Journal. 34: 574-80.

Georgelin-Gurgel M, Devillard R, Lauret ME, Diemer F, Calas P, Hennequin M. 2008. Root canal shaping using rotary nickel-titanium files in preclinical teaching. Odonto-stomatologie Tropicale. 31:5–11.

Haapasalo M, Shen Y. 2013. Evolution of nickel–titanium instruments: from past to future. Endodontic Topics;29:3–17.

Hänni S, Schönenberger K, Peters OA, Barbakow F. 2003. Teaching an engine-driven preparation technique to undergraduates: initial observations. International Endodontic Journal. 36:476-82. DOI 10.1046/j.1365-2591.2003.00677.x.

Hieawy A, Haapasalo M, Zhou H, Wang ZJ, Shen Y. 2015. Phase Transformation Behavior and Resistance to Bending and Cyclic Fatigue of ProTaper Gold and ProTaper Universal Instruments. Journal of Endodontics.41:1134-8. DOI 10.1016/j.joen.2015.02.030.

Kang MS, Kim HC, Hur B, Park JK. 2006. Comparison Of Shaping Ability of Rotary Ni-Ti File Systems Used by Undergraduates. Restorative Dentistry & Endodontics.; 31:1-10. DOI 10.5395/JKACD.2006.31.1.001.

- Kum KY, Spängberg L, Cha BY, Il-Young J, Seung-Jong L, Chan-Young L. 2000. Shaping ability of three ProFile rotary instrumentation techniques in simulated resin root canals. *Journal of Endodontics*.;26:719–23. DOI 10.1097/00004770-200012000-00013.
- Kwak SW, Cheung GS, Ha JH, Kim SK, Lee H, Kim HC. 2016. Preference of undergraduate students after first experience on nickel-titanium endodontic instruments. *Restor Dent Endod*. 41:176-81. DOI 10.5395/rde.2016.41.3.176.
- Lambrianidis T. 2009. Ledging and blockage of rootcanals during canal preparation:causes, recognition, prevention,management, and outcomes. *Endodontic Topics* 2009, 15, 56–7474.
- Martins RC1, Seijo MO, Ferreira EF, Paiva SM, Ribeiro Sobrinho AP. 2012. Dental students' perceptions about the endodontic treatments performed using NiTi rotary instruments and hand stainless steel files. *Brazilian Dental Journal*. 23:729-36. DOI 10.1590/S0103-64402012000600018.
- Nagy CD, Bartha K, Bernáth M, Verdes E, Szabó J. 1997.The effect of root canal morphology on canal shape following instrumentation using different techniques. *International Endodontic Journal*. 30:133–40. DOI 10.1046/j.1365-2591.1997.00049.x
- Parashos P, Messer HH. 2006. Rotary NiTi instrument fracture and its consequences. *Journal of Endodontics*. 32:1031-43. DOI 10.1016/j.joen.2006.06.008

- 420 Pérez-Higueras JJ, Arias A, de la Macorra JC, Peters OA. 2014. Differences in cyclic
421 fatigue resistance between ProTaper Next and ProTaper Universal instruments
422 at different levels. *Journal of Endodontics*. 40:1477-81. DOI
423 10.1016/j.joen.2014.02.025.
- 424 Peru M, Peru C, Mannocci F, Sherriff M, Buchanan LS, Pitt Ford TR. 2006. Hand and
425 nickel-titanium root canal instrumentation performed by dental students: a micro-
426 computed tomographic study. *European Journal of Dental Education*. 10:52-9.
427 DOI 10.1111/j.1600-0579.2006.00395.x
- 428 Peters CI, Schönenberger K, Barbakow F. 2003. ProTaper rotary root canal
429 preparation: effects of canal anatomy on final shape analysed by micro CT.
430 *International Endodontic Journal*. 36:86-92. DOI 10.1046/j.1365-2591.
- 431 .Peters OA, Boessler C, Zehnder M. 2005. Effect of liquid and paste-type lubricants on
432 torque values during simulated rotary root canal instrumentation. *International*
433 *Endodontic Journal*. DOI 38:223–9. 10.1111/j.1365-2591.2005.00937.x
- 434 Qualtrough AJ, Dummer PM. 1997. Undergraduate endodontic teaching in the United
435 Kingdom: an update. *International Endodontic Journal*. 30: 234-9. DOI
436 10.1046/j.1365-2591.1997.t01-1-00072.x.
- 437 Ribeiro DM, Réus JC, Felipe WT, Pacheco-Pereira C, Dutra KL, Santos JN Porporatti,
438 AL, De Luca Canto G. 2018. Technical quality of root canal treatment performed
439 by undergraduate students using hand instrumentation: a meta-analysis.
440 *International Endodontic Journal*.51:269-283. doi: 10.1111/iej.12853.

- 441 Saunders WP, Saunders EM. 1992. Effect of noncutting tipped instruments on the
- 442 quality of root canal preparation using a modified double-flared technique.
- 443 Journal of Endodontics. 18: 32-6. DOI 10.1016/S0099-2399(06)81140-4.
- 444 Schäfer E, Vlassis M. 2004. Comparative investigation of two rotary nickel-titanium
- 445 instruments: ProTaper versus RaCe. Part 1. Shaping ability in simulated curved
- 446 canals. International Endodontic Journal. 37: 239-48. DOI 10.1111/j.0143-
- 447 2885.2004.00783.x.
- 448 Schneider SW. 1971. A comparison of canal preparations in straight and curved root
- 449 canals. Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology.32:271-
- 450 5.
- 451 Sonntag D, Bärwald R, Hülsmann M, Stachniss V. 2008. Pre-clinical endodontics: a
- 452 survey amongst German dental schools. International Endodontic Journal.
- 453 41:863-8. DOI 10.1111/j.1365-2591.2008.01438.x.
- 454 Thompson SA. 2000. An overview of nickel-titanium alloys used in dentistry.
- 455 International Endodontic Journal. 33:297-310. DOI 10.1046/j.1365-
- 456 2591.2000.00339.x
- 457 Tu MG1, Chen SY, Huang HL, Tsai CC. 2008. Endodontic shaping performance using
- 458 nickel-titanium hand and motor ProTaper systems by novice dental students.
- 459 Journal of the Formosan Medical Association. 107:381-8. DOI 10.1016/S0929-
- 460 6646(08)60103-5.
- 461 Tulsa Dental Specialties. Protaper Universal. Directions for use (2017)
- 462 [http://www.tulsadentalspecialties.com/Libraries/Tab_Content-](http://www.tulsadentalspecialties.com/Libraries/Tab_Content-Endo_Access_Shaping/DFUPTNF_Rev1_10-ProTaperNext_DFU.sflb.ashx)
- 463 [Endo_Access_Shaping/DFUPTNF_Rev1_10-ProTaperNext_DFU.sflb.ashx](http://www.tulsadentalspecialties.com/Libraries/Tab_Content-Endo_Access_Shaping/DFUPTNF_Rev1_10-ProTaperNext_DFU.sflb.ashx) .

Unal GC, Kececi AD, Kaya BU, Tac AG. 2011. Quality of root canal fillings performed by undergraduate dental students. *European Journal of Dentistry*. 5:324-30.

Unal GC, Maden M, Orhan EO, Saritekin E, Teke A. 2012. Root canal shaping using rotary nickel-titanium files in preclinical dental education in Turkey. *Journal of Dental Education*. 76:509-13.

Unal G.C, Maden M, Savgat A, Orhan EO. 2009 "Comparative investigation of 2 rotary nickel-titanium instruments: protaper universal versus protaper", *Oral Surgery Oral Medicine Oral Pathology Oral Radiology and Endodontology* (ISI), 886-892 pp. DOI 10.1016/j.tripleo.2009.01.010.

Vaudt J, Bitter K, Neumann K, Kielbassa AM. 2009. Ex vivo study on root canal instrumentation of two rotary nickel-titanium systems in comparison to stainless steel hand instruments. *International Endodontic Journal*. 42:22-33. DOI 10.1111/j.1365-2591.2008.01489.x.

Walia HM, Brantley WA, Gerstein H. 1988. An initial investigation of the bending and torsional properties of Nitinol root canal files. *Journal of Endodontics*;14:346–51. DOI 10.1016/s0099-2399(88)80196-1.

Vieira EP, França EC, Martins RC, Buono VT, Bahia MG.. 2008. Influence of multiple clinical use on fatigue resistance of ProTaper rotary nickel-titanium instruments. *International Endodontic Journal*. 41:163–72. DOI 10.1111/j.1365-2591.2007.01336.x.

Yang GB, Zhou XD, Zheng YL, Zhang H, Shu Y, Wu HK. 2007. Shaping ability of progressive versus constant taper instruments in curved root canals of extracted

teeth. International Endodontic Journal. 40: 707-14. DOI 10.1111/j.1365-
2591.2007.01296.x.

Yang G, Yuan G, Yun X, Zhou X, Liu B, Wu H. 2011. Effects of two nickel-titanium
instrument systems, Mtwo versus ProTaper universal, on root canal geometry
assessed by micro-computed tomography. Journal of Endodontics. 37:1412-6.
DOI 10.1016/j.joen.2011.06.024.

Zehnder M. 2006. Root canal irrigants. Journal of Endodontics. 32:389 –98. DOI
10.1016/j.joen.2005.09.014.

Zuolo M L, Walton R E. Instrument deterioration with usage: nickel-titanium versus
stainless steel. Quintessence Int 1997; 28: 397–402.

Table 1(on next page)

The canal curvature, the change in WL, the working time, the fractured instrument and deformed instrument, and the ledge based on the groups after instrumentation

SD: standard deviation, PTU: ProTaper Universal, PTN: Protaper Next, N: the number of sample

	PTU	N	PTN	N	total	P value
The degree and SD of canal curvature	24.03°±3.14°	174	25.64°±2.72°	176	395	0.000***
Amount of change and SD in WL (mm)	0.97±0.95	155	0.96±0.80	174	395	0.99
Working time and SD (s)	627±28	200	379±18	195	395	0.000
The number and percent of the canal with fractured instrument	17.4 % (n=33)	190	9.32 % (n=18)	194	395	0.020*
The percent of deformed instrument	4.5 % n=9	200	2.6% n=2	78	278	0.403
The percent of ledge	18.8 % n=33	176	4% n=7	176	352	0.000***

1

2

Figure 1

Measurement of the canal curvature on the images captured by stereo microscope after instrumentation

a: degree of curvature according to Schneider (1971)




Figure 2

Ledge formation at the beginning of the curvature associated with a marked loss of working length

The arrow indicates the ledge.

